

CERTIFICATE ELECTROMAGNETIC COMPATIBILITY

Applicant : **Mart Microbiology B.V.**
Contact person : **Mr. P. Helder**
Address : **Galileïstraat 26**
Postal code, Place : **7130 AD, Lichtenvoorde**
Country : **The Netherlands**

Manufacturer : **Mart Microbiology B.V.**
Address : **Galileïstraat 26**
Postal code, Place : **7131 PE, Lichtenvoorde**
Country : **The Netherlands**


Electrical apparatus : **Instrument for the cultivation of micro-organisms in an oxygen-depleted atmosphere**
Trademark : **Mart Microbiology**
Type designation : **Anoxomat Mark II, AN2CTS**

Environment : **Commercial, light industrial and industrial**

EN 61326:1998 Equipment for measurement, control and laboratory use, from which:
EN 55011:1998 Emission - Class A
EN 61000-3-2:1995 Harmonic current emissions
EN 61000-4-2:1995 Electrostatic discharge (ESD) immunity
EN 61000-4-3:1998 Radiated Electro-Magnetic field immunity
ENV 50204:1995 Digital radio telephones immunity
EN 61000-4-4:1995 Electrical fast transient (EFT) immunity
EN 61000-4-5:1995 Surge transient immunity
EN 61000-4-6:1996 Conducted Radio-Frequency disturbances immunity
EN 61000-4-11:1994 Immunity to voltage variations, dips and short interrupts

The undersigned declares that the described products meet the essential requirements of the **EMC Directive 89/336/EEC**, based on a non-recurrent examination.
The results are recorded in our test report with reference 2037038-QUA/EMC 03-4848.

KEMA Quality B.V.
(Competent Body EMC)
Arnhem, 29 October, 2003



A.T. van der Meijden
Certification Manager EMC

Certificate nr. **2037038-QUA/EMC 03-4849**