



IEC SYSTEM FOR CONFORMITY
TESTING TO STANDARDS FOR
SAFETY OF ELECTRICAL
EQUIPMENT (IECEE) CB SCHEME

Ref. Certificate No.

NL-7472

CB TEST CERTIFICATE

Issued by: KEMA Quality B.V.

Registration date: 2004-02-04

Product: Anoxomat Mark II

Applicant: Mart Microbiology B.V. Galileïstraat 26 The Netherlands
7131 PE Lichtenvoorde

Manufacturer: Mart Microbiology B.V. Galileïstraat 26 The Netherlands
7131 PE Lichtenvoorde

Factory: Mart Microbiology B.V. Galileïstraat 26 The Netherlands
7131 PE Lichtenvoorde

Rating and principal characteristics: 220-240 V, 50-60 Hz, 3,4 A or
110-120 V, 50-60 Hz, 3,4 A

Trade mark (if any): MART

Model/Type reference: AN2CTS

Additional information:

Sample of product tested to be in conformity with IEC: 61010-1(ed.2)

Test Report Ref. No: 2037038.50-QUA/MED

This CB Test Certificate is issued by the National Certification Body

KEMA Quality B.V.
Utrechtseweg 310
P.O. Box 5185
6802 ED Arnhem
The Netherlands



Signed by: B.B. van Leeuwen

Date of issue: 2004-02-04